

# 九十一年度 G6PD 院際品管檢體活性測定的再現性

## 低活性品管檢體G6PD活性測定的再現性

91.02 ~ 91.10

檢驗單位	n	Mean	S.D.	C.V.(%)
RH-01	6	2.1	0.5	22.2
RH-02	6	3.0	0.2	7.7
RH-03	6	1.9	0.6	32.6
RH-04	6	3.8	0.4	9.4
RH-05	6	2.9	0.8	26.2
RH-06	6	2.5	0.3	13.4
RH-07	6	2.6	0.4	14.2
RH-08	6	2.1	0.2	10.3
RH-09	6	2.8	0.3	11.7
RH-10	6	2.4	0.5	22.1
RH-11	6	2.4	0.7	29.5
RH-12	6	2.6	0.2	7.0
RH-13	6	3.0	0.5	16.9
RH-14 <sup>1</sup>	6	4.1	4.1	100.1
RH-16	1	1.0	--	--
RH-17	6	2.3	0.9	41.2
RH-18	6	2.8	0.4	15.7
RH-19	6	2.6	0.4	16.3
A0189	6	2.6	0.7	26.6
A0203	6	2.5	0.4	17.5
G026	6	2.8	0.2	7.0
G031	6	2.3	0.5	20.3
參考值 <sup>2</sup>	6	2.7	0.15	5.7

\* 送檢批號：9102，9104，9106，9107，9110

- 1、 RH-14扣除一次異常結果，Mean = 3.0，S.D.= 0.4，CV = 13.3%
- 2、 Determined at Clin. Biochem. Res. Lab., Taipei Veterans General Hospital

## 高活性品管檢体G6PD活性測定的再現性

91.02 ~ 91.10

檢驗單位	n	Mean	S.D.	C.V.(%)
RH-01	4	9.7	1.4	14.4
RH-02	4	10.2	0.9	8.7
RH-03	4	10.5	1.1	10.6
RH-04 <sup>1</sup>	4	12.2	1.6	13.0
RH-05	4	11.5	0.4	3.9
RH-06	4	9.9	0.7	7.5
RH-07	4	10.4	1.0	9.4
RH-08	4	9.8	0.8	7.7
RH-09	4	10.8	1.1	9.7
RH-10 <sup>1</sup>	4	10.4	1.3	12.5
RH-11	4	9.0	0.9	10.4
RH-12	4	10.7	0.6	5.8
RH-13	4	10.1	0.7	7.0
RH-14 <sup>2</sup>	4	8.1	3.7	46.2
RH-16	0	--	--	--
RH-17	4	9.4	1.1	11.4
RH-18	4	11.6	2.0	17.3
RH-19	4	10.3	0.6	6.0
A0189	4	10.3	0.4	3.7
A0203	4	11.0	2.0	18.1
G026	4	11.2	1.2	10.6
G031	4	9.7	1.4	14.2
參考值 <sup>3</sup>	4	10.4	0.5	4.8

\* 送檢批號：9102，9106，9107

- 1、 RH-04扣除一次異常結果，Mean = 11.5，S.D.= 1.1，CV = 9.4%
- 2、 RH-14扣除一次異常結果，Mean = 10.0，S.D.= 0.3，CV = 2.9%
- 3、 Determined at Clin. Biochem. Res. Lab., Taipei Veterans General Hospital