

九十七年度 G6PD 院際品管檢體活性測定的再現性

低活性品管檢體 G6PD 活性測定的再現性

96.01 ~ 97.11

檢驗單位	n ¹	Mean	S.D.	C.V.(%)
RH-01	5	5.4	0.48	8.9%
RH-02	5	4.9	0.35	7.1%
RH-03	5	5.0	0.81	16.2%
RH-04	5	4.8	0.68	14.2%
RH-06	5	5.0	0.45	9.0%
RH-07	5	4.7	0.94	20.0%
RH-08	5	5.3	0.43	8.1%
RH-09	5	5.0	0.44	8.8%
RH-10	5	5.4	0.53	9.8%
RH-12	5	5.3	0.53	10.0%
RH-13	5	4.9	0.64	13.1%
RH-14	5	5.1	0.61	12.0%
RH-18	5	4.8	0.74	15.4%
RH-19	5	5.1	0.28	5.5%
A0189	5	4.4	0.89	20.2%
A0203	5	5.0	0.62	12.4%
B001	5	5.4	0.74	13.7%
G026	5	5.3	0.17	3.2%
G031	5	5.3	0.53	10.0%
CL001	5	4.9	0.32	6.5%
CL002	5	4.7	0.66	14.0%
參考值 ²	5	5.1	0.14	2.7%

1、 送檢批號：RH2007-03、RH2007-07、RH2007-09、RH2008-03、RH2008-11

2、 Determined at Clin. Biochem. Res. Lab., Taipei Veterans General Hospital

中活性品管檢體 G6PD 活性測定的再現性

96.01 ~ 97.11

檢驗單位	n ¹	Mean	S.D.	C.V.(%)
RH-01	5	8.3	0.29	3.5%
RH-02	5	7.8	0.30	3.8%
RH-03	5	7.7	0.51	6.6%
RH-04	5	7.5	0.33	4.4%
RH-06	5	8.2	0.51	6.2%
RH-07	5	8.1	0.78	9.6%
RH-08	5	8.4	0.61	7.3%
RH-09	5	7.8	0.69	8.8%
RH-10	5	8.3	0.40	4.8%
RH-12	5	7.9	1.03	13.0%
RH-13	5	8.1	0.89	11.0%
RH-14	5	8.3	0.90	10.8%
RH-18	5	7.6	0.74	9.7%
RH-19	5	8.3	0.45	5.4%
A0189	5	7.9	0.86	10.9%
A0203	5	7.6	0.40	5.3%
B001	5	8.4	1.07	12.7%
G026	5	8.8	1.08	12.3%
G031	5	9.0	0.82	9.1%
CL001	5	8.0	0.35	4.4%
CL002	5	7.8	0.89	11.4%
參考值 ²	5	8.0	0.05	0.6%

1、 送檢批號：RH2007-05、RH2007-07、RH2007-11、RH2008-01、RH2008-07

2、 Determined at Clin. Biochem. Res. Lab., Taipei Veterans General Hospital

高活性品管檢體 G6PD 活性測定的再現性

96.01 ~ 97.11

檢驗單位	n ¹	Mean	S.D.	C.V.(%)
RH-01	5	13.3	0.91	6.8%
RH-02	5	12.8	0.39	3.0%
RH-03	5	12.5	0.90	7.2%
RH-04	5	12.4	0.43	3.5%
RH-06	5	13.2	0.83	6.3%
RH-07	5	13.2	0.86	6.5%
RH-08	5	13.6	0.48	3.5%
RH-09	5	12.5	0.62	5.0%
RH-10	5	13.3	0.86	6.5%
RH-12	5	12.5	1.50	12.0%
RH-13	5	13.6	1.13	8.3%
RH-14	5	13.1	1.77	13.5%
RH-18	5	12.2	0.78	6.4%
RH-19	5	13.1	0.85	6.5%
A0189	5	13.0	1.34	10.3%
A0203	5	12.4	0.86	6.9%
B001	5	12.7	0.83	6.5%
G026	5	14.1	1.37	9.7%
G031	5	13.7	1.19	8.7%
CL001	5	12.5	0.62	5.0%
CL002	5	13.2	0.91	6.9%
參考值 ²	5	12.7	0.08	0.6%

1、 送檢批號：RH2007-01、RH2007-07、RH2007-11、RH2008-01、RH2008-07

2、 Determined at Clin. Biochem. Res. Lab., Taipei Veterans General Hospital